



All dimensions are in mm; tolerances according to ISO 2768 m-H

**Interface**

According to IEC 61169-40  
Mechanically compatible with RPC-1.85

**Contents and Documentation**

- This kit is delivered with
- **Standard Definitions Card**  
Printed Standard Definitions that can be used on nearly all Vector Network Analyzers
  - **Test Results Documentation**
  - **Hard Shell Case**
  - **Protection Caps**

**Material and plating**

**Connector parts**

Center conductor  
Outer conductor  
Body  
Dielectric  
Substrate

**Material**

CuBe  
Stainless steel  
Aluminum  
PS  
Al<sub>2</sub>O<sub>3</sub>

**Plating**

Gold, min. 1.27 µm, over nickel  
Passivated  
black anodized

**Electrical data**

Frequency range	DC to 50.0 GHz
<b>Thru</b>	
Return loss	≥ 30 dB, DC to 4 GHz ≥ 24 dB, 4 GHz to 26.5 GHz ≥ 17 dB, 26.5 GHz to 50 GHz
<b>Open</b>	
Error from nominal phase <sup>1</sup>	≤ 2.0°, DC to 4 GHz ≤ 4.0°, 4 GHz to 26.5 GHz ≤ 6.0°, 26.5 GHz to 50.0 GHz
<b>Short</b>	
Error from nominal phase <sup>2</sup>	≤ 1.5°, DC to 4 GHz ≤ 3.0°, 4 GHz to 26.5 GHz ≤ 4.5°, 26.5 GHz to 50.0 GHz
<b>Load</b>	
Return loss	≥ 36.0 dB, DC to 4 GHz ≥ 30.0 dB, 4 GHz to 26.5 GHz ≥ 22.0 dB, 26.5 GHz to 50.0 GHz
DC Resistance	50 Ω ± 0.5 Ω
Power handling (at 25 °C, sea level)	≤ 0.5 W, derated linearity by 0.005 W/K

<sup>1</sup> The nominal phase is defined by the Offset Delay, the Offset Loss and the Fringing Capacitances

<sup>2</sup> The nominal phase is defined by the Offset Delay, the Offset Loss and the Short Inductance

**Mechanical data**

Mating cycles	≥ 500
Maximum torque	1.65 Nm
Recommended torque	0.90 Nm
Gauge	0.00 mm to 0.05 mm

**General standard definitions**

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

**Thru**

Offset Z <sub>0</sub> / Impedance / Z <sub>0</sub>	50 Ω
Offset Delay	87.394 ps
Length (electrical) / Offset Length	26.20 mm
Offset Loss	4.00 GΩ/s
Loss	0.0304 dB/√GHz
Line Loss @ 1GHz	0.0012 dB/mm

**Open**

Offset Z <sub>0</sub> / Impedance / Z <sub>0</sub>	50 Ω
Offset Delay	23.350 ps
Length (electrical) / Offset Length	7.00 mm
Offset Loss	3.20 GΩ/s
Loss	0.0130 dB/√GHz
Fringing Capacitances	C <sub>0</sub> = 4.30000 x 10 <sup>-15</sup> F / 4.30000 fF
	C <sub>1</sub> = -718.000 x 10 <sup>-27</sup> F/Hz / -0.71800 fF /GHz
	C <sub>2</sub> = 28.7000 x 10 <sup>-36</sup> F/Hz <sup>2</sup> / 0.02870 fF /GHz <sup>2</sup>
	C <sub>3</sub> = -0.30000 x 10 <sup>-45</sup> F/Hz <sup>3</sup> / -0.00030 fF /GHz <sup>3</sup>

**Short**

Offset $Z_0$ / Impedance / $Z_0$	50 $\Omega$	
Offset Delay	23.350 ps	
Length (electrical) / Offset Length	7.00 mm	
Offset Loss	3.50 G $\Omega$ /s	
Loss	0.0142 dB/ $\sqrt{\text{GHz}}$	
Short Inductance	$L_0 = 4.00000 \times 10^{-12} \text{ H}$	/ 4.00000 pH
	$L_1 = 0.00000 \times 10^{-24} \text{ H/Hz}$	/ 0.00000 pH/GHz
	$L_2 = 0.00000 \times 10^{-33} \text{ H/Hz}^2$	/ 0.00000 pH/GHz <sup>2</sup>
	$L_3 = 0.00000 \times 10^{-42} \text{ H/Hz}^3$	/ 0.00000 pH/GHz <sup>3</sup>

**Load**

Offset $Z_0$ / Impedance / $Z_0$	50 $\Omega$
Offset Delay	0.0000 ps
Length (electrical) / Offset Length	0.000 mm
Offset Loss	0.00 G $\Omega$ /s
Loss	0.0000 dB/ $\sqrt{\text{GHz}}$

**Environmental data**

Operating temperature range <sup>3</sup>	+20 °C to +26 °C
Rated temperature range of use <sup>4</sup>	0 °C to +50 °C
Storage temperature range	-40 °C to +85 °C
RoHS	compliant

<sup>3</sup> Temperature range over which these specifications are valid.

<sup>4</sup> This range is underneath and above the operating temperature range, within the calibration kit is fully functional and could be used without damage

**Declaration of documentation**

Standard delivery for this kit includes Test Results. The documentation issued reports which quantities were tested individually, traceable to national / international standards. Model based standard definitions of the calibration standards are reported in Agilent / Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

**Inspection interval**

Recommendation 12 months

**Packing**

Standard 1 pce in bag  
Weight 36.3 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
Marcel Panicke	06.09.18	Markus Müller	01.02.19	c00	19-0220	Marion Striegler	01.02.19